

Specification Sheet

Specification Number: Revision: Date Issued: 8688J-330 01 22-Jul-19

1. Part Number:

8688J-330

2. Photograph:



3. Product Information:

Description:	SteriWare [®] Jumbo Sack Sampler - 325mm
Use:	Single use, Sack Sampler
Material of Construction:	 HDPE (High Density Polyethylene) - virgin Conforms to FDA 21 CFR 177.1520 Conforms to EU Regulations 10/2011 Conforms to EC Regulations 1935/2004 Conforms to EC Regulations 2023/2006 BSE/TSE free
Method of Construction:	Single piece, injection moulded
Overall Length:	325mm
Overall Diameter:	45mm
Max Internal Diameter:	36mm
Beak Length:	95mm
Beak Width:	36mm
Nominal Weight:	61g
Moulding & Primary Packing Environment:	Class 100,000 Cleanroom
Individually Bagged?	Yes (heat sealed PE/PET bag)
Method of Sterilisation:	N/A
Number of Samplers per Box:	100
Recommended Storage Conditions:	Dry and ambient temperature
Shelf Life:	5 years from the date of manufacture
ATEX:	This device has not been ATEX tested



4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK.

4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



Specification Sheet Specification Number: 8688J-330 Revision: Date Issued: 22-Jul-19

01

4. Change Control:

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to change.control@sampling.com

This is a single-use, disposable sampler. Repeated use will lead to

5. Country of Origin:

Made and packed in the UK.

6. Repeated Use:

7. Approval:

Signed:

failure.

Kuldeep Flora Name: Position: **Quality Manager** Date: 22-Jul-19



Sampling Systems Ltd. 4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com